



FAPAS QC MATERIAL DATA SHEET	T06179QC
Matrix	Cod Liver Oil
Weight / Volume of Contents	30 ml

Analyte	Assigned Value, X_a	Range for $ z \leq 2$	Units	No. of data points producing X_a
BDE 28	0.310	0.174 - 0.446	$\mu\text{g}/\text{kg}$	3
BDE 47	1.10	0.62 - 1.58	$\mu\text{g}/\text{kg}$	3
BDE 49	0.100	0.056 - 0.144	$\mu\text{g}/\text{kg}$	3
BDE 99	1.15	0.64 - 1.66	$\mu\text{g}/\text{kg}$	3
BDE 100	0.800	0.448 - 1.152	$\mu\text{g}/\text{kg}$	3
BDE 153	1.58	0.88 - 2.28	$\mu\text{g}/\text{kg}$	3
BDE 154	0.640	0.358 - 0.922	$\mu\text{g}/\text{kg}$	3
Sum of PBDE excluding BDE-209	5.69	3.19 - 8.19	$\mu\text{g}/\text{kg}$	3

This data sheet is applicable until	13 Sep 2028
Recommended Storage on receipt	Ambient
Notes	
<ul style="list-style-type: none">• Mix the QC material thoroughly before taking a representative analytical sample• The assigned value has been derived from the consensus of laboratories taking part in this proficiency test, using a variety of methods. This is not a certified reference value.• The Range for $z \leq 2$ is the concentration range within the limits of ± 2 z-scores. The assigned value and its range have been established from the proficiency test data and are suitable for use by laboratories as a fit-for-purpose quality control measure.• Stability of the QC material has been established as sufficient for the scope of the proficiency test from previous experience, expert advice and published literature. FAPAS advises that the QC material is analysed within the recommended date. FAPAS QC materials are intended to be used as single-analysis samples.• Full details on the proficiency test procedure used to characterise this QC material are available in the Protocol, Part 1 - Common Principles, freely available to download from the FAPAS website.• You may use any method of analysis you wish.	